



VitreK QT Enterprise 95X Test Report

Page 1

DUT Model Your Model

DUT Serial Brief Format

Tested By: operator on 4/4/2022 at 14:34

Using Model 952i S/N 017909 with F/W v2.39

PASSED ALL TESTS

Step #1 DCW @ 100.00V for 2.00s PASSED
Step #2 ACW @ 200.0V/60.00Hz for 2.00s PASSED
Step #3 DCIR @ 200.0V for 2.00s PASSED
Step #4 GB @ 10.00A/60.00Hz for 2.00s PASSED
Step #5 Low Ohms PASSED
Step #6 AC Gnd Leakage for 2.00s PASSED
Step #7 DC Gnd Leakage for 2.00s PASSED
Step #8 Breakdown at 10.00mA PASSED
Step #9 Positive Pulse at 100V PASSED